AquaPūrST

Technical Data Sheet

– BRIEF DESCRIPTION –

AquaPur™ST is Purified Water that has been sterilized and packaged in a double bag for use in the cleanroom areas within pharmaceutical, biotechnology, and medical device facilities or anywhere the use of sterile, USP Purified water is required.

AquaPur™ST is prepared and assayed to specifications which meet USP Purified Water Reagent Grade Water. AquaPur™ST meets CAP and NCCLS Standards for Reagent Grade Type 1 water at time of manufacture. AquaPur™ST contains no preservatives. It is gamma-irradiated to a sterility assurance level (SAL) of 10^{-6}. Each lot is tested for Endotoxin via USP LAL testing and sterility via USP 14-day sterility test. A copy of lot specific documentation, which details conductivity, TOC, irradiation, endotoxin, sterility, and water quality limits for USP Purified Water, is attached to every case of AquaPur™ ST.

- PRECAUTIONARY STATEMENT -

Not to be used for Parenteral Administration. Not for use in sterile irrigation, injection, infusion, or as a potable product. AquaPur™ST is intended for manufacturing and general laboratory use.

- STORAGE -

Store in cool area. Sterility can not be assured if the plastic bags are not sealed. AquaPur™ST has an expiration date of two years and the date is listed on the bottle.

- PRODUCT SPECIFICATIONS -

AquaPur™ Sterile, USP grade Purified Water, is filtered through a 0.2 µm filter. *AquaPur™ ST* is tested to meet USP specifications for Purified Water.

USP Water Conductivity USP <645>	≤ 1.3 umhos/cm
USP Total Organic Carbon USP <643>	< 0.5 PPM

- PACKAGING -

1 Gallon (3.8 Liters) / 4 per case

- CATALOG NUMBER -

Manufactured for:	Item	Cat. No.
Decon Labs, Inc.	**1 gallon**	**8904**

For Additional Technical Information call
Decon Customer Service
(800) 332-6647, or
See our website at www.deconlabs.com

®Decon Labs, Inc. *Revised 8/17/2011*

460 Glennie Circle • King of Prussia, PA 19406 • (610) 755-0800 • Fax (610) 270-8905